

SHIJIAZHUANG HONGRAY GROUP

South Tongda Rd., East Dist. Jinzhou City
Hebei, 052260, China
TEL: 86-311-66179662
www.hongray.com

SPECIFICATION FOR NITRILE EXAMINATION GLOVE

1.0 Product: Nitrile, Powder Free, Blue, 9"length, for examination use, NPF3001-3006

2.0 Dimensions:

Size	Median Length (mm)	Median Width (mm)	Thickness (mm)(min)	
			Palm	Finger
XS	240	70 ± 10	0.06	0.07
S	240	80 ± 10	0.06	0.07
M	240	95 ± 10	0.06	0.07
L	240	110 ± 10	0.06	0.07
XL	240	120 ± 10	0.06	0.07
Note: Thickness for reference only, not as a basis for rejection.				

3.0 Strength:

Force at break: 6N (Median)

4.0 Water Leakage Testing: According to ISO2859, G-I, Single Sampling Plan, AQL1.0

5.0 Powder Residues: Not more than 2mg per glove

DOCUMENT NO.: NBR-SETINO-003E/001

INITIALED DATE: 2022.10.08

REVISION DATE: N/A

REVISION TIMES: 0