

# SHIJIAZHUANG HONGRAY GROUP

South Tongda Rd., East Dist. Jinzhou City  
Hebei, 052260, China  
TEL: 86-311-66179662  
www.hongray.com

## SPECIFICATION FOR NITRILE EXAMINATION GLOVE

**1.0 Product:** Nitrile, Powder Free, Black, 9"length, for examination use, NPF4001-4006

### 2.0 Dimensions:

Size	Median Length (mm)	Median Width (mm)	Thickness (mm)(min)	
			Palm	Finger
XS	240	70±10	0.09	0.11
S	240	80±10	0.09	0.11
M	240	95±10	0.09	0.11
L	240	110±10	0.09	0.11
XL	240	120±10	0.09	0.11
Note: Thickness for reference only, not as a basis for rejection.				

### 3.0 Strength:

Force at break: 6N (Median)

**4.0 Water Leakage Testing:** According to ISO2859, G-I, Single Sampling Plan, AQL1.5

**5.0 Powder Residues:** Not more than 2mg per glove

DOCUMENT NO.: NBR-SETINO-004E

INITIALED DATE: 2022.10.08

REVISION DATE: N/A

REVISION TIMES: 0