

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	1 of 24

TECHNICAL DOCUMENTATION

(acc. to MDR (EU) 2017/745 Annex II & III)

Product: Vinyl Examination Gloves, Powder Free, Non-Sterile

Product Group Code: VGPF100X

Revision: 01

Date: October 8, 2022

Prepared by: QA Department

Reviewed by: Wu min

Approved by: Zhang Yaping

Electronic file is valid without a signature

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	2 of 24

TABLE OF CONTENTS

#	CONTENTS	PAGES
1.0	Device description and specification, including variants and accessories	3
2.0	Information to be supplied by the manufacturer	11
3.0	Design and manufacturing information	12
4.0	General safety and performance requirements	18
5.0	Benefit-risk analysis and risk management	18
6.0	Product verification and validation	18
7.0	Post-market surveillance	21
8.0	Reference	24

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	3 of 24

TECHNICAL DOCUMENTATION

1. DEVICE DESCRIPTION AND SPECIFICATION, INCLUDING VARIANTS AND ACCESSORIES

1.1 Device description and specification

(a) Product or trade name and a general description of the device including its intended purpose and intended users

Product or Trade Name

Vinyl Examination Gloves, Powder Free, Non-Sterile

Product Group Code

VGPF100X

Legal Manufacturer

Shijiazhuang Hongray Group Co., Ltd.

South Tongda Rd., East Dist. Jinzhou City, Hebei, 052260, China

Manufacturing Facilities

Site 1: Grand Work Plastic Products Co., Ltd

Donggao Industrial Zone, Zhanhuang, Hebei, 050000, China

Site 2: Ever Light Plastic Products Co., Ltd

Donggao Industrial Zone, Zhanhuang, Shijiazhuang, Hebei, 050000, China

Site 3: Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County, Linfen, Shanxi, 042300, China

Site 4: Shijiazhuang Hongjun Plastic Manufacture Co., Ltd.

Xinye Industrial Zone, Wangtong Village, Wancheng Town, Gaoyi County, Shijiazhuang, Hebei, 051330, China

European Authorised Representative

Caretechion GmbH

Address: Niederrheinstr. 71, 40474 Duesseldorf, Germany

Contact: Mr Wang


Tel: +49 211 3003 6618

Email: info@caretechion.de

Table 1: General Description

NO.	Topic	Description
1	Product name	Vinyl Examination Gloves, Powder Free, Non-Sterile

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	4 of 24

2	Product group code	VGPF100X
3	Product code	VGPF1001- VGPF1006
4	Product specification code	PVC-SETINO-001E
5	Type	Single-use disposable examination glove
6	Marking	All information on dispenser box
7	Shape	Ambidextrous
8	Material	Vinyl
9	Cuff/ surface	Cuff Beaded / Textured at fingertip or Full Textured
10	Color	Clear
11	Available sizes	Extra Small (XS) Small (S) Medium (M) Large (L) Extra Large (XL)
12	Photographs	

Intended Purpose

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Vinyl Examination Gloves, Powder Free, Non-Sterile are intended for medical activities except surgery.

Intended Users

Medical personnel, who perform medical examinations, diagnostic and therapeutic procedures, as well as personnel, who work with contaminated medical materials.

(b) The Basic UDI-DI as referred to in Part C of Annex VI assigned by the manufacturer to the device in question, as soon as identification of this device becomes based on a UDI system, or otherwise a clear identification by means of product code, catalogue number or other unambiguous reference allowing traceability

The Basic UDI-DI of Vinyl Examination Gloves, Powder Free, Non-Sterile must be placed on packaging after 05/26/2025.

The definition is as follows:

XXXXXXXX: Company Prefix registered number with GS1 represents for Shijiazhuang Hongray Group Co., Ltd.

XXXX: Product Reference as Vinyl Examination Gloves, Powder Free, Non-Sterile

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	5 of 24

XX: Check Character that has been calculated and validated on GS1 website

(c) The intended patient population and medical conditions to be diagnosed, treated and/or monitored and other considerations such as patient selection criteria, indications, contraindications, warnings

Intended Patient Population

The intended patient population is for medical personnel, who perform medical examinations, diagnostic and therapeutic procedures, as well as personnel, who work with contaminated medical materials.

Examination gloves are intended for all kind of patients, with the requirement of hygienic medical therapy and/or examination, both superficial skin and natural body orifices. It is not intended for patients, that require surgical procedures. Respective sensitizations and allergy warnings must be respected.

Medical Conditions to be Diagnosed, Treated and/or Monitored

The examination gloves are used in medical or healthcare facilities indoors in a dry place, away from direct sunlight, at recommended temperature.

The examination gloves are intended for medical activities except surgery.

Indications

The indications for use of the medical device are the necessity of conduction of medical examinations, diagnostic and therapeutic procedures.

Table 2: Summary of the Indications for Gloving and for Glove Removal

	Indication
Gloves on	1) Before hygienic procedures 2) When anticipating contact with blood or another body fluid, regardless of the existence of sterile conditions and including contact with non-intact skin and mucous membrane 3) Contact with a patient (and his/her immediate surroundings) during contact precautions
Gloves off	1) As soon as gloves are damaged (or non-integrity suspected) 2) When contact with blood, another body fluid, non-intact skin and mucous membrane has occurred and has ended 3) When contact with a single patient and his/her surroundings, or a contaminated body site on a patient has ended 4) When there is an indication for hand hygiene 5) When healthcare facility internal guidelines require the removal of the glove

Contraindications/ Warning

The contraindications/ warning of the examination gloves are as follows:

- For single use only;
- For medical examination;
- It is necessary to perform hygienic treatment of your hands before putting on or changing gloves, as well as after removing the gloves;
- It is necessary to choose the gloves of the right size;

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	6 of 24

- It is necessary to change the gloves in case of any defect, puncture, damage, contamination, etc.;
- It is necessary to put the gloves on dry hands only;
- Keep in dry condition and avoid the sunlight;

(d) Principles of operation of the device and its mode of action, scientifically demonstrated if necessary

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Vinyl Examination Gloves, Powder Free, Non-Sterile are intended for medical activities except surgery.

The operational principles might vary, depending on healthcare facilities and operations done with the product. The usage procedure is self-explanatory and further instructed during the training of healthcare personnel. The general contradictions and warnings apply to all procedures.

(e) The rationale for the qualification of the product as a device

Medical gloves are used to prevent the transmission of disease and provide a safety barrier against contamination. They are typically used during examinations or medical procedures. As such, they are medical devices.

(f) The risk class of the device and the justification for the classification rule(s) applied in accordance with Annex VIII, Examination Gloves, Powder Free, Non-Sterile are medical device, class I.

The product is classified in accordance with Rule 1 and Rule 5 of Annex VIII:

“Rule 1: All non-invasive devices are classified as class I, unless one of the rules set out hereinafter applies”

“Rule 5: All invasive devices with respect to body orifices, other than surgically invasive devices, which are not intended for connection to an active device or which are intended for connection to a class I active device are classified as: Class I if they are intended for transient use”.

The rationale for classification is shown as table below.

Table 3: Classification Rationale

Classification Rules Acc. to Annex VIII	Applicable (Y/N)	Rationale
Non-invasive devices, Rule 1	Y	The examination gloves are disposable device intended for medical purposes that are worn on the examiners hand or finger to prevent contamination between patient and examiner. Examination gloves are intended for medical activities except surgery and mostly used for external, so it could be determined that the device is non-invasive as class I.
Non-invasive devices, Rule 2	N	The examination gloves are not intended for channelling or storing blood, body liquids, cells or tissues, liquids or gases for the purpose of eventual infusion, administration or introduction into the body.
Non-invasive devices, Rule 3	N	The examination gloves are not intended for modifying the biological or chemical composition of human tissues or cells, blood, other body liquids or other liquids intended for implantation or administration into the body. Also the examination gloves are not consisted of a substance or a mixture of substances intended to be used in vitro.

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	7 of 24

Non-invasive devices, Rule 4	N	The examination gloves are not intended for come into contact with injured skin or mucous membrane.
Invasive devices, Rule 5	Y	The examination gloves are also intended for transient use in term of invasive device with respect to body orifices such as gynecologist or dental work.
Invasive devices, Rule 6 - 8	N	The examination gloves are intended for medical activities except surgery, so the device is not classified as surgically invasive devices as well as implantable devices.
Active devices, Rule 9 - 13	N	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner, so the device is not classified as active devices.
Special rule 14	N	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner, so the device is not classified as medicinal product.
Special rule 15	N	The examination gloves are not intended to use for contraception or prevention of the transmission of sexually transmitted diseases.
Special rule 16	N	The examination gloves are not intended specifically to be used for disinfecting, cleaning, rinsing or, where appropriate, hydrating contact lenses. Also the examination gloves are not intended specifically to be used for disinfecting or sterilizing medical devices.
Special rule 17	N	The examination gloves are not intended specifically for recording of diagnostic images generated by X-ray radiation.
Special rule 18	N	The examination gloves are not manufactured utilising tissues or cells of human or animal origin, or their derivatives, which are non-viable or rendered non-viable.
Special rule 19	N	The examination gloves are not incorporating or consisting of nanomaterial.
Special rule 20	N	The examination gloves are not intended as invasive devices with respect to body orifices to administer medicinal products by inhalation.
Special rule 21	N	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner. The examination gloves are intended for medical activities except surgery, so the device is not composed of substances or of combinations of substances that are intended to be introduced into the human body via a body orifice or applied to the skin and that are absorbed by or locally dispersed in the human body.
Special rule 22	N	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner, so the device is not classified as active therapeutic devices.

(g) An explanation of any novel features

This requirement is not applicable, as examination gloves are not claimed as novel devices.

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	8 of 24

(h) A description of the accessories for a device, other devices and other products that are not devices, which are intended to be used in combination with it

This requirement is not applicable, as examination gloves are not worn with other accessories or used in combination.

(i) A description or complete list of the various configurations/variants of the device that are intended to be made available on the market

The examination gloves will be packed and made available on the market under form of its packaging.

The package provides protection from the exposure to mechanical and climatic factors during transportation and storage. The package consists of primary and secondary packaging as follows;

Primary packaging or Dispenser box: The gloves are packed into dispenser box. The dispenser box provides protection of gloves from external mechanical damage and makes it possible to determine visually the identification and traceability of the gloves.

Secondary packaging or Transport carton: The dispenser boxes are packed into transport carton. Transport carton boxes are strong enough and provide safety of devices during transportation and storage.

Therefore, a description or complete list of the various configurations/variants of the device that are intended to be made available on the market is shown as table below.

Table 4: Complete List of the Various Configurations/Variants of the Device



NO.	Topic	Description
1	Glove sizes	<p>Available sizes</p> <p>Extra Small (XS)</p> <p>Small (S)</p> <p>Medium (M)</p> <p>Large (L)</p> <p>Extra Large (XL)</p> <p>Dimension: Palm width</p> <p>XS: < 80 mm</p> <p>S: 80 ± 10 mm</p> <p>M: 95 ± 10 mm</p> <p>L: 110 ± 10 mm</p> <p>XL: > 110 mm</p> <p>Dimension: length</p> <p>Median 240 mm for all sizes</p> <p>Dimension: Single wall thickness, for reference only</p> <p>Detail refer to Specification</p>
2	Glove packing sizes	<p>The quantity of packing/packages shall follow the agreed specification for each brand, according to Approval Packaging & Labels that could be variant quantity as the following samples;</p> <ul style="list-style-type: none"> Gloves are packed 100 pcs into each dispenser box, then 10 dispenser boxes are packed into each transport carton

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	9 of 24

3	Packaging sizes	Pack 100 pcs per each dispenser box, then pack 10 dispenser boxes per each transport carton. The dimension will be as actual need or customer's requirements.
4	Packaging materials	Dispenser box: Paper greyback Transport carton: Brown corrugated

(j) A general description of the key functional elements, e.g. its parts/components (including software if appropriate), its formulation, its composition, its functionality and, where relevant, its qualitative and quantitative composition. Where appropriate, this shall include labelled pictorial representations (e.g. diagrams, photographs, and drawings), clearly indicating key parts/components, including sufficient explanation to understand the drawings and diagrams

Table 5: General Description of the Key Functional Elements

NO.	Topic	Description
1	General description of the key functional elements	Vinyl Examination Gloves, Powder Free, Non-Sterile are made of Vinyl and chemicals. The gloves are online dipping PU solution. The PU reduces the surface tackiness of the gloves allowing for gloves to be easily donned without powder.
2	Photographs	
3	Drawing	 Designation of length and width of gloves

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	10 of 24

The list of materials/compositions used for the manufacturing of Vinyl Examination Gloves, Powder Free, Non-Sterile is shown in table below:

Table 6: Material Used During Manufacturing of the Product

PVC resin

Plasticizer, (DOTP, DINP)

Stabilizer

TXIB

PU releasing agent

Pigment (for gloves with colors)

(k) A description of the raw materials incorporated into key functional elements and those making either direct contact with the human body or indirect contact with the body, e.g., during extracorporeal circulation of body fluids
Refer to Table 6: Material Used During Manufacturing of the Product.

(l) Technical specifications, such as features, dimensions and performance attributes, of the device and any variants/configurations and accessories that would typically appear in the product specification made available to the user, for example in brochures, catalogues and similar publications

Table 7: Technical Specifications (Quality Characteristics)

Description	Specification	Test-Method
BARRIER PROPERTIES Freedom from holes	AQL \leq 1.5	EN 455-1
Powder residue on powder free gloves	\leq 2.0 mg/glove	EN ISO 21171 Monitoring follows EN 455-3
PHYSICAL PROPERTIES Force at break (during shelf life)	EN 455-2 Median 3.6 N	EN 455-2
DIMENSION Hand-width size related	EN 455-2 Size related table issued on request XS: < 80 mm S: 80 ± 10 mm M: 95 ± 10 mm L: 110 ± 10 mm XL: > 110 mm	EN 455-2
Total length	EN 455-2 Median 240 mm for all sizes	EN 455-2

Performance Requirements for Quality Characteristics

- The sampling plan for in-process and final inspection follows in accordance with ISO 2859-1 “Sampling procedures for inspection by attributes – Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection;
- The Acceptance Quality Level shall conform to the AQL mentioned in agreed specification document
- The product meets the provision of MDR (EU) 2017/745 Class I and the latest revision of EN 455 all parts.

1.2 Reference to previous and similar generations of the device

(a) An overview of the previous generation or generations of the device produced by the manufacturer, where such

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	11 of 24

devices exist

N/A

(b) An overview of identified similar devices available on the Union or international markets, where such devices exist

The equivalent devices of Vinyl Examination Gloves products. Non-Sterile Examination Gloves by Blue sail. Examination Gloves have already been on the market for years.

Vinyl Glove offers the best combination of extreme friendly comfort and furious reliable hand protection (EN455) while improving the sustainability of Vinyl gloves.

2. INFORMATION TO BE SUPPLIED BY THE MANUFACTURER

(a) The label or labels on its packaging, such as inner box and transport carton / outer case in case of specific management conditions, in the languages accepted in the Member States where the device is envisaged to be sold.

The package labeling and marking of symbols will be shown on packaging of product as follows;

- Dispenser box: this is a primary packaging that the gloves will be packed into a dispenser box;
- Transport carton: this is a secondary packaging that the dispenser boxes above will be packed into a transport carton

The package labeling and marking of symbols shall comply with the agreed specification of packaging for each brand according to Approval Packaging & Labels as well as in accordance with the applicable regulations/ standards such as:

- EN 1041 Information supplied by the manufacturer of medical devices;
- EN ISO 15223-1 Symbols to be used with medical device labels, labelling and information to be supplied;
- EN 455-3 Requirements and testing for biological evaluation (item – labelling)

(Reference: procedure refers to Packaging and Labeling Control)

(b) The instructions for use in the languages accepted in the Member States where the device is envisaged to be sold

The requirement for the instruction for use is not applicable for class I devices as the device can be used safely without any such instructions, however the recommended instruction for use could be determined as below.

Instruction for Use

1. Always wash hands before putting on gloves and each time you change to a new pair.
2. Take the gloves out from their original box.
3. Select glove appropriate for the task and in your size. Inform yourself prior to the application.
4. Touch only a restricted surface of the glove corresponding to wrist (at the top edge of the cuff).
5. Don the first glove by inserting your hand into the opening and pulling it over by applying a soft pull by the second hand.
6. Take the second glove with bare hand and touch only a restricted surface of the glove corresponding to wrist.
7. To avoid touching the skin of the forearm with the gloved hand, turn the external surface of the glove to be donned on the folded fingers of the gloved hand so there is a good hold on it. Thus, permitting to glove the second hand by pulling softly with the first gloved hand.
8. Once gloved, hands should not touch anything else that is not defined by indications and conditions for glove use.

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	12 of 24

9. Change gloves before beginning a different task to avoid cross contamination.

10. Pinch one glove at the wrist level to remove it, without touching the skin of the forearm, and peel away from the hand, thus allowing the glove to turn inside out.

11. Hold the removed glove in the gloved hand and slide the fingers of the ungloved hand inside between the glove and the wrist of the gloved hand. Remove the second glove by rolling it down the hand and fold the first glove into the outside palm area of the second glove. If done correctly, the user holds a pack of the outer glove turned inside and out and the contaminated surface of both gloves is within the package.

12. Discharge the removed gloves and wash hands thoroughly according the respective hygiene standards.

The shelf life is normally advised to be three full years. Check the original dispenser instructions for details.

3. DESIGN AND MANUFACTURING INFORMATION

(a) For Vinyl Examination Gloves: According to the requirements of device classification and conformity evaluation in ANNEX VIII of the Medical Device Regulations (MDR)2017/745/EU, examination gloves are Class I medical devices. According to this regulation, low-risk medical devices below Class II can be implemented in accordance with Appendix XI PART A, which does not contain design controls.

Vinyl Examination Gloves are mature and marketed products, our company has not made any design changes.

(b) Complete information and specifications, including the manufacturing processes and their validation, their adjuvants, the continuous monitoring, and the final product testing. Data shall be fully included in the technical documentation

Table 8: Complete Information and Specifications

Description	Specification	Test-Method
BARRIER PROPERTIES Freedom from holes	$AQL \leq 1.5$	EN 455-1
Powder residue on powder free gloves	≤ 2.0 mg/glove	EN ISO 21171 Monitoring follows EN 455-3
PHYSICAL PROPERTIES Force at break (during shelf life)	EN 455-2 Median 3.6 N	EN 455-2
DIMENSION Hand-width size related	EN 455-2 Size related table issued on request XS: < 80 mm S: 80 ± 10 mm M: 95 ± 10 mm L: 110 ± 10 mm XL: > 110 mm	EN 455-2
Total length	EN 455-2 Median 240 mm for all sizes	EN 455-2

Performance Requirements for Quality Characteristics

The sampling plan for in-process and final inspection follows in accordance with ISO 2859-1 "Sampling procedures

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	13 of 24


for inspection by attributes – Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection;

The Acceptance Quality Level shall conform to the AQL mentioned in agreed specification document

The product meets the provision of MDR (EU) 2017/745 Class I and the latest revision of EN 455 all parts.

Manufacturing Process

The manufacturing process for Vinyl Examination Gloves, Powder Free, Non-Sterile is shown as below:



Validation

The validation for Vinyl Examination Gloves, Powder Free, Non-Sterile consists of validation Protocol and validation report which shown as follows:

Validation Protocol

(Performance Qualification Protocol of Vinyl Examination Gloves, Powder Free, Non-Sterile,

1. Purpose

To establish confidence through appropriate testing that the finished product produced by process of group line Vinyl Examination Gloves, Powder Free, Non-Sterile are effective and meet all release requirements for functionality and safety.

2. Scope

This Performance Qualification Protocol is applicable for group line of production process of Vinyl Examination Gloves, Powder Free, Non-Sterile.

3. Process Description and Principle of Operation

According to manufacturing process flow chart, work instruction of production process for Vinyl Examination

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	14 of 24

Gloves, Powder Free, Non-Sterile.

4. Requirements for the Process

According to parameter specifications of production process for Vinyl Examination Gloves, Powder Free, Non-Sterile.

5. Inspection/ Test Method for Measuring the Evaluation Results

According to In-Process and Final Inspection Sampling Plan.

6. Manufacturing Materials, Equipment and Calibration

To ensure the approval of materials, availability of equipment and calibration status.

7. Acceptance Criteria

The performance qualification for group line of production process of Vinyl Examination Gloves, Powder Free, Non-Sterile will be acceptable if fulfilment of following criteria;

- All process parameters meet specification defined.
- All properties of product meet on requirements.

8. Conclusion

To summarize the conclusion of validation for Performance Qualification of Vinyl Examination Gloves, Powder Free, Non-Sterile.

9. Established by

Assistant Technical Assurance Manager

10. Reviewed and Approved by

Assessment and Technical Assurance Manager

Validation Report

(Performance Qualification Report of Vinyl Examination Gloves, Powder Free, Non-Sterile)

1. Purpose

To establish confidence through appropriate testing that the finished product produced by process of group line Vinyl Examination Gloves, Powder Free, Non-Sterile are effective and meet all release requirements for functionality and safety.

2. Scope

This Performance Qualification Protocol is applicable for group line of production process of Vinyl Examination Gloves, Powder Free, Non-Sterile.

3. Process Description and Principle of Operation

According to manufacturing process flow chart, work instruction of production process for Vinyl Examination Gloves, Powder Free, Non-Sterile.

4. Requirements for the Process

According to parameter specifications of production process for Vinyl Examination Gloves, Powder Free, Non-Sterile.

5. Inspection/ Test Method for Measuring the Evaluation Results

According to In-Process and Final Inspection Sampling Plan.

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	15 of 24

6. Manufacturing Materials, Equipment and Calibration

The approval of materials, availability of equipment and calibration status, are completed prior validation.

7. Acceptance Criteria

The performance qualification for group line of production process of Vinyl Examination Gloves, Powder Free, Non-Sterile will be acceptable if fulfilment of following criteria;

- All process parameters meet specification defined.
- All properties of product meet on requirements.

8. Conclusion

The performance qualification for group line of production process of Vinyl Examination Gloves, Powder Free, Non-Sterile is completed, acceptable and complied to the following criteria;

- All process parameters meet specifications defined.
- All properties of product meet on requirements.

9. Established by

Assistant Quality Assurance Manager

10. Reviewed and Approved by

Quality Assurance Manager

Quality Assurance Monitoring and the Final Product Testing


Table 11: Quality Assurance Monitoring and the Final Product Testing

No.	QA Inspection and Testing	Description
1	Incoming materials inspection	<p>This step is to inspect and ensure that the quality of materials include Vinyl, chemicals, packaging and formers meet the specifications defined of each material.</p> <pre> graph TD A[Receiving of materials] --> B{Testing/ Inspection and Evaluation} B -- Pass --> C[Storage] B -- Fail --> D[Concession] B -- Fail --> E[Reject] E -- "Complaint to Supplier" --> A </pre> <p>Flow chart of incoming materials inspection</p>

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	16 of 24

2	Ingredient and compounding preparation inspection	<p>This step is to inspect and ensure that the quality of ingredients and compounds after preparation meet the specifications defined before approval to use in production line.</p> <pre> graph TD A[Chemical prepared by IP&CP] --> B{LAB Test} B -- Pass --> C[Use in production line] B -- Fail --> D[Concession] B -- Fail --> E[Rework] D --> F[Use in production line] E --> B </pre> <p>Flow chart of ingredient and compounding preparation inspection</p>
3	Production line/ chemical in line inspection	<p>This step is to check and ensure that parameters control in production line meet the specifications defined as well as to inspect and ensure that the chemicals in production line meet the specifications defined.</p> <pre> graph TD A[Chemical use in production line] --> B{LAB Test} B -- Pass --> C[Use in production line] B -- Fail --> D[Retest] B -- Fail --> E[Rework] D --> F[Use in production line] E --> B </pre> <p>Flow chart of chemical in line inspection</p>

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	17 of 24



(c) Identification of all sites, including suppliers and sub-contractors, where manufacturing activities are performed.

Manufacturing Site Facilities

Site 1: Grand Work Plastic Products Co., Ltd

Donggao Industrial Zone, Zanzhuang, Hebei, 050000, China

Site 2: Ever Light Plastic Products Co., Ltd

Donggao Industrial Zone, Zanzhuang, Shijiazhuang, Hebei, 050000, China

Site 3: Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County, Linfen, Shanxi, 042300, China

Site 4: Shijiazhuang Hongjun Plastic Manufacture Co., Ltd.

Xinye Industrial Zone, Wangtong Village, Wancheng Town, Gaoyi County, Shijiazhuang, Hebei, 051330, China

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	18 of 24

4. GENERAL SAFETY AND PERFORMANCE REQUIREMENTS

- (a) The general safety and performance requirements that apply to the device and an explanation as to why others do not apply
- (b) The method or methods used to demonstrate conformity with each applicable general safety and performance requirement
- (c) The harmonized standards or other solutions applied
- (d) The precise identity of the controlled documents offering evidence of conformity with each harmonized standard, or other method applied to demonstrate conformity with the general safety and performance requirements. The information referred to under this point shall incorporate a cross-reference to the location of such evidence within the full technical documentation and, if applicable, the summary technical documentation
- (Reference: General Safety and Performance Requirements Checklist, Vinyl Powder Free Examination Gloves, Non-Sterile)

5. BENEFIT-RISK ANALYSIS AND RISK MANAGEMENT

- (a) The benefit-risk analysis referred to in Sections 1 and 8 of Annex I

Based on the conducted risk analysis for the Vinyl Examination Gloves, Powder Free, Non-Sterile product, the foreseeable risks have been identified and evaluated in most cases as acceptable with respect to the intended application and use of the products. Counteractions have been taken for those items for which an initially unacceptable risk has been identified. Subsequently, the performed implementations were verified. Finally, it can be determined that no unacceptable residual risk exists with the products either individually or cumulatively, that outweighs the benefits from the use of the products.

- (b) The solutions adopted and the results of the risk management referred to in Section 3 of Annex I

Based on the conducted risk analysis for the Vinyl Examination Gloves, Powder Free, Non-Sterile product, the foreseeable risks have been identified and evaluated in most cases as acceptable with respect to the intended application and use of the products. Counteractions have been taken for those items for which an initially unacceptable risk has been identified. Subsequently, the performed implementations were verified. Finally, it can be determined that no unacceptable residual risk exists with the products either individually or cumulatively, that outweighs the benefits from the use of the products.

(Reference: Risk Management Plan and Risk Analysis Report of Vinyl Powder Free Examination Gloves, Non-Sterile)

6. PRODUCT VERIFICATION AND VALIDATION

6.1. Pre-Clinical and Clinical Data

- (a) Results of tests, such as engineering, laboratory, simulated use and animal tests, and evaluation of published literature applicable to the device, taking into account its intended purpose, or to similar devices, regarding the pre-clinical safety of the device and its conformity with the specifications
- (b) Detailed information regarding test design, complete test or study protocols, methods of data analysis, in addition

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	19 of 24

to data summaries and test conclusions regarding in particular

- The biocompatibility of the device including the identification of all materials in direct or indirect contact with the patient or user and the tests carried out on the gloves showed no indication of a potential cause for skin irritation and sensitization as well as Cytotoxicity

The testing reports above mentioned are available upon request.

- Physical, chemical and microbiological characterization

The physical, chemical and microbiological characterization tests for Vinyl Examination Gloves shown as the following table:

#	Test Items	Reference Standards	Summary Result
1	Freedom from holes	EN 455-1	Pass
2	Dimension	EN 455-2	Pass
3	Force at Break	EN 455-2	Pass
4	Residual Powder	EN 455-3	Pass

Test report is available upon request.

- Electrical safety and electromagnetic compatibility

This requirement is not applicable, as examination gloves are not relating to electrical safety and electromagnetic compatibility.

- Software verification and validation

This requirement is not applicable, as software is not used in finished devices – examination gloves.

- Stability/Shelf Life

Based on conducted the stability study according to EN455-4 in-house testing, the shelf life for Vinyl Examination Gloves, Powder Free, Non-Sterile is 3 years

Performance and safety

Usage Validation

The design of this product has been validated to ensure the performance and safety of examination gloves according to the following evaluation items:

- Fit/ comfort
- Skin irritation
- Tactile sensitivity (dry, wet)
- Grip
- Ease of donning
- Odor
- Color
- Strength/ durability

Summary: the result of usage validation is acceptable.

(c) The clinical evaluation report and its updates and the clinical evaluation plan referred to in Article 61(12) and Part A of Annex XIV.

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	20 of 24

Conclusion of Clinical Evaluation

The clinical evaluation clearly demonstrates that the Vinyl Examination Gloves, Powder Free, Non-Sterile are in conformity with the relevant general safety and performance requirements of MDR 2017/745 (Annex I) by compliance to the requirements of EN 455. The current information material is considered adequate with respect to the intended user.

The performance and safety of Vinyl Examination Gloves, Powder Free, Non-Sterile have been established and risk associated with the use of these devices are acceptable when weighed against the benefits to the patient.

Taken together, the performance and safety of the devices have been established and risks associated with the use of the devices are acceptable when weighed against the benefit for the user and patient.

(d) The PMCF plan and PMCF evaluation report referred to in Part B of Annex XIV or a justification why a PMCF is not applicable

The PMCF Plan outlines the methods and chosen frames for updating the clinical evaluation on a regular basis, incorporating new sources of information and checking on identified risks and the correct assessment of those in the risk management activities.

Examination gloves are used since centuries and the materials PVC have been well established with several hundred billion gloves per year consumed all over the world. From this fact and the current post-market surveillance data available, there might be the possibility explored, that PMCF studies are not necessary to evaluate the safety, quality or performance of that particular medical devices.

6.2. Additional information required in specific cases

(a) Where a device incorporates, as an integral part, a substance which, if used separately, may be considered to be a medicinal product within the meaning of point 2 of Article 1 of Directive 2001/83/EC, including a medicinal product derived from human blood or human plasma, as referred to in the first subparagraph of Article 1(8), a statement indicating this fact. In this case, the documentation shall identify the source of that substance and contain the data of the tests conducted to assess its safety, quality, and usefulness, taking account of the intended purpose of the device.

This requirement is not applicable, as examination gloves are not classified as medicinal product devices.

(b) Where a device is manufactured utilizing tissues or cells of human or animal origin, or their derivatives, and is covered by this Regulation in accordance with points (f) and (g) of Article 1 (6, and where a device incorporates, as an integral part, tissues or cells of human origin or their derivatives that have an action ancillary to that of the device and is covered by this Regulation in accordance with the first subparagraph of Article 1 (10), a statement indicating this fact. In such a case, the documentation shall identify all materials of human or animal origin used and provide detailed information concerning the conformity with Sections 13.1. or 13.2., respectively, of Annex I.

This requirement is not applicable, as examination gloves are not manufactured utilizing tissues or cells of human or

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	21 of 24

animal origin, or their derivatives.

(c) In the case of devices that are composed of substances or combinations of substances that are intended to be introduced into the human body and that are absorbed by or locally dispersed in the human body, detailed information, including test design, complete test or study protocols, methods of data analysis, and data summaries and test conclusions

This requirement is not applicable, as examination gloves are not composed of substances or combinations of substances that are intended to be introduced into the human body and that are absorbed by or locally dispersed in the human body.

(d) In the case of devices containing CMR or endocrine-disrupting substances referred to in Section 10.4.1 of Annex I, the justification referred to in Section 10.4.2 of that Annex

This requirement is not applicable, as examination gloves do not contain of CMR or endocrine-disrupting substances.

(e) In the case of devices placed on the market in a sterile or defined microbiological condition, a description of the environmental conditions for the relevant manufacturing steps. In the case of devices placed on the market in a sterile condition, a description of the methods used, including the validation reports, with respect to packaging, sterilization, and maintenance of sterility. The validation report shall address bioburden testing, pyrogen testing and, if applicable, testing for sterilant residues

This requirement is not applicable, as this kind of examination gloves is non-sterile and do not define microbiological condition.

(f) In the case of devices placed on the market with a measuring function, a description of the methods used in order to ensure the accuracy as given in the specifications

This requirement is not applicable, as examination gloves are not classified as measuring function devices.

(g) If the device is to be connected to other device(s) in order to operate as intended, a description of this combination/configuration including proof that it conforms to the general safety and performance requirements when connected to any such device(s) having regard to the characteristics specified by the manufacturer

This requirement is not applicable, as examination gloves are not to be connected to other devices in order to operate as intended.

7.0 Post-market surveillance

7.1. The post-market surveillance plan drawn up in accordance with Article 84

Post-Market Surveillance Plan

The post-market surveillance (PMS) is the activities carried out to gain information about the quality, safety or

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	22 of 24

performance of medical devices that have been placed in the market.

The post-market surveillance and action system consists of four main steps as follows:

- **Step 1 Surveillance Input**

The surveillance input is a channel to gain information about the quality, safety or performance of medical devices that have been placed in the market, which could be determined as the table below.

Plan & Time Schedule

NO.	Action	Timeframe	Desired Outcome
1	Complaint analysis	Without undue delay after incoming complaint information	Proper case studies and statistics
2	Competitor analysis	At least once a year	Comparison & studying of advancement of state of the art
3	Literature review	At least all two years	Proper literature section in the clinical evaluation report
4	Interviews with users, customers, conferences, regulators and notified bodies	During trade shows, fairs, customer visits and inspections	Advanced understanding of changed settings, applications, experiences, complaints, technologies and regulations
5	Review of authority reports & actions taken by regulators	Without undue delay after receiving an authority reports or learning about an action taken by a regulatory	Learnings from case studies
6	Review of public incident reporting databases	Once a year	Learnings from case studies
7	Preclinical Data & Testing	Per Demand	Preclinical assessments shall help to study product performances under standardized and objective settings.
8	Proactive questionnaires	At least once a year	Raising the level of understanding and awareness on customer side & getting feedback on more complex and rare product deficiencies

- **Step 2 Investigation & Analysis**

After receiving the post-market surveillance information above, the person in charge and all involved in each PMS topic above shall perform investigate and analysis those information as well as to find and summarize the investigation and analysis output for further action and communication.

- **Step 3 Action**

The action shall be taken following the output from investigation and analysis in order to monitor and maintain the

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	23 of 24

quality, safety or performance of medical devices. Include the action shall comply with the applicable standards, regulations and laws in each market.

Example of action; response of complaint investigation, adverse event reporting, advisory notice issuing, recall/correction and removal, corrective action and prevention action; process, labeling, training, and so on.

● Step 4 Communication

Together with action shall be considered for communication also. The communication shall comply with the applicable standards, regulations and laws in each market.

Example of communication; to communicate to Stakeholders, Regulatory Authorities, Representative, Notified Body, Customers, Distributors, Suppliers and so on.

The overall meeting for post-market surveillance (PMS) will be performed at least once a year to summarize the overall feedback through the year in order to ensure the products are still under safety, quality and performance of medical devices.

7.2. The PSUR referred to in Article 86 and the post-market surveillance report referred to in Article 85

Regarding the post-market surveillance, Hongray Group has appropriate procedures in place to handle feedback from the market, the last overall meeting for post-market surveillance (PMS) could be determined the output from a meeting as follows.

Post-Market Surveillance Report

NO.	Action	Timeframe	Report
1	Complaint analysis	Without undue delay after incoming complaint information	Proper case studies and statistics “The determined complaint rates reside in a general accepted level, there is no critical complaints concerning alleged deficiencies related to safety or performance that require to recall or withdrawal the products from the market, there is no competent authority post market surveillance inquiries”
2	Competitor analysis	At least once a year	Comparison & studying of advancement of state of the art “The determined competitor analysis shows the comparison and studying of competitors for improvement process”
3	Literature review	At least all two years	Proper literature section in the clinical evaluation report “The determined literature reviews show the studying of literature for improvement process and proper literature section in the clinical evaluation report”
4	Interviews with users, customers, conferences, regulators and notified bodies	During trade shows, fairs, customer visits and inspections	Advanced understanding of changed settings, applications, experiences, complaints, technologies and regulations “The determined interviews with users, customers, conferences, regulators and notified bodies show advanced understanding of changed settings, applications, experiences, complaints, technologies and regulations”
5	Review of authority reports & actions taken by regulators	Without undue delay after receiving an authority reports or learning about an action taken by a regulatory	Learnings from case studies “The determined review of authority reports & actions taken by regulators show the learnings from case studies for improvement process”

TECHNICAL FILE MDR	Document Number	TF-MDR-VGPF-01
	Issue Date	October 8, 2022
	Revision	01
	Page	24 of 24

NO.	Action	Timeframe	Report
6	Review of public incident reporting databases	Once a year	Learnings from case studies “The determined review of public incident reporting databases shows the learnings from case studies for improvement process”
7	Preclinical Data & Testing	Per Demand	Preclinical assessments shall help to study product performances under standardized and objective settings. “The determined preclinical data & testing shows all product performances meet the standardized and objective settings”
8	Proactive questionnaires	Once a year	Raising the level of understanding and awareness on customer side & getting feedback on more complex and rare product deficiencies

Therefore, it could be determined that the products are still under safety, quality and performance of medical devices.

8.0 REFERENCES

List of Applicable Regulations and Standards

No	Regulation/ Standard Number	Regulation/ Standard Name
1	MDR (EU) 2017/745	Medical Device Regulation
2	EN ISO 13485: 2016	Medical devices - Quality management systems - Requirements for regulatory purposes
3	EN ISO 9001: 2015	Quality management systems – requirements
4	ISO 14971: 2019	Medical devices - application of risk management to medical devices
5	EN 455-1: 2020/A1:2022	Requirements and testing for freedom from holes
6	EN 455-2: 2015	Requirements and testing for physical properties
7	EN 455-3: 2015	Requirements and testing for biological evaluation
8	EN 455-4: 2009	Requirements and testing for shelf life determination
9	ISO 10993-1: 2020	Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process
10	ISO 10993-5: 2009	Biological evaluation of medical devices – Part 5: Test for in vitro cytotoxicity
11	ISO 10993-10: 2010	Biological evaluation of medical devices – Part 10: Test for irritation and skin sensitization
12	EN 1041: 2008+A1: 2013	Information supplied by the manufacturer of medical devices
13	EN ISO 15223-1: 2016	EN ISO 15223-1 Symbols to be used with medical device labels, labelling and information to be supplied
14	EN 16523-1: 2015+A1: 2018	Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact