

SHIJIAZHUANG HONGRAY GROUP

South Tongda Rd., East Dist. Jinzhou City
Hebei, 052260, China
TEL: 86-311-66179662
www.hongray.com

SPECIFICATION FOR VINYL EXAMINATION GLOVE

1.0 Product: Vinyl, Powder Free, Transparent, 9"length, for examination use, VGPF 1001-1006

2.0 Dimensions:

Size	Median Length (mm)	Median Width (mm)	Thickness (mm)(min)	
			Palm	Finger
XS	240	75±5	0.07	0.07
S	240	85±5	0.07	0.07
M	240	95±5	0.07	0.07
L	240	105±5	0.07	0.07
XL	240	115±5	0.07	0.07
Note: Thickness for reference only, not as a basis for rejection.				

3.0 Strength:

Force at break: 3.6N (Median)

4.0 Water Leakage Testing: According to ISO2859, G-I, Single Sampling Plan, AQL1.5

5.0 Powder Residues: Not more than 2mg per glove

DOCUMENT NO.: PVC-SETINO-001E

INITIALED DATE: 2022.10.08

REVISION DATE: N/A

REVISION TIMES: 0